



CENTRAL SERICULTURAL GERMPLOSM RESOURCES CENTRE
CENTRAL SILK BOARD

MINISTRY OF TEXTILES, GOVT. OF INDIA
P.B. NO.44, THALLY ROAD, HOSUR - 635 109, KRISHNAGIRI DIST., TAMIL NADU.

Phone: (91) 4344 222013, 221148 Fax: (91) 4344 220520

E-mail: csgrchsr@dataone.in, director_csgrc@yahoo.com Website: www.silkgermplasm.com

9

No.CSB/GRC/3(3)/2012/STORES/DBT/S.W./ 1982 to 1991. Date:6th February, 2012

To:

10 Nos. suppliers, as per list enclosed (individually addressed)

Sirs,

Sub:- Inviting competitive rate quotations for supply of Electroporator
to CSGRC, CSB, Hosur - reg.

This office is interested in purchase of 1 No. Electroporator as per the detailed specification, enclosed. You are requested to quote your rate duly indicating the discount to be allowed on quoted value and Tax/Transportation charge, if any. You may also enclose Catalogue, users list and performance report from the users, Warranty, Authorization & validity certificate from the manufacturer.

Quotation should be sent to the undersigned in a sealed envelope super scribing as the "QUOTATION FOR SUPPLY OF 1 No. ELECTROPORATOR", so as to reach this office on or before 27-2-2012 at 2.30 p.m. The quotations received after the due date and Time of receipt will be rejected. The Quotations will be opened by a Committee on the same day at 3.00 p.m.

The undersigned reserves the right to accept or reject any or all quotations without assigning any reason whatsoever.

The Terms and Conditions are as under:-

- (1) EMD @ 2# on the quoted rate in the form of Demand Draft favoring "The Director, Central Sericultural Germplasm Resources Centre" payable at Hosur. EMD will be refunded only after finalization of the bids, without interest.
- (2) Validity : Rate quoted should be valid for 90 days.
- (3) Delivery : Within 15 days from the date of receipt of the Purchase order.
- (4) Rate per unit should be quoted such as per piece / per metre etc. & Destination: CSGRC, Hosur. Payment schedule : By Demand Draft or cheque after receipt of the stores in good condition and conforming to specification.
- (5) Taxes/VAT/Other statutory duties should be indicated separately. Please note that no form "C" or "D" will be issued by this station.
- (6) Liquidated damages for delay & non supply: 0.5% will be deducted from the bill per week for not supplying the work within the stipulated time.
- (7) Submission of income tax clearance as issued by Income Tax Department & sales Tax clearance certificate as issued by State / Central Sales tax Department.

Yours faithfully,

DIRECTOR

Encl: Specification

sar.

Specification for Electroporator

- Fast and freely programmable-two program buttons allow storage and recall of most commonly used parameters, default settings : P1 1700V and P2 2500V
- Programmable voltage and optimized fixed pulse times
- Safe electronics and integrated electroporation chamber eliminate voltage leaks and misuse.
- Easy, PLP-complaint documentation via USB port-data
- Integrated cuvette holder
- Display should set parameters and actual voltage and time constant after the pulse
- Input voltage : 100 V ~ 240 V \pm 10 %, 50 Hz-60 Hz
- Power input : 20 w
- Pulse voltage: 200 – 2500 V
- Out put wavelength : Decaying exponential waveform
- Time constant : 5 ms
- Charging time : < 10 s
- Resistance : 600 Ω
- Capacity : 100 μ F
- Interface : USB 2.0
- Dimension (LxWxH) : 27.5 x 19 x 12.5 cm
- Weight: 3.2 kg